The X5CAM4K Series HDMI/NETWORK/USB Multi-outputs C-mount CMOS Camera Help Manual



## Contents

Con	tents			I			
1	X5C	AM4K Ser	ies Camera Application	1			
2	X5CAM4K Series Camera Datasheet and Functions (2)						
3	Dim	Dimension of X5CAM4K Series Camera					
4	X5CA	AM4K Ser	ies Camera Packing Information	5			
5	Soft	ware and	Арр	7			
6	X5CA	AM4K Ser	ies Camera Configurations	8			
	6.1	Camer	a working standalone with built-in XCamView software	8			
	6.2	Conne	cting camera to computers with USB3.0 port	9			
	6.3	Camer	a working in WiFi mode (AP mode)	10			
	6.4	Conne	cting camera to the PC with LAN port	12			
	6.5 appl	Conne ication	cting multi-cameras to the router through the LAN port/ WiFi STA mode for the net	<i>w</i> ork 13			
7	Brief	fIntroduc	tion of X5CAM4K UI and Its Functions	17			
	7.1	XCam\	/iew UI	17			
	7.2	The ca	mera control panel on the left or right side of the video window	17			
	7.3	The M	easurement Toolbar on top of the video window	18			
	7.4	Icons a	and functions of the Synthesis Camera Control Toolbar at the bottom of the video window	19			
		7.4.1	Browse	19			
		7.4.2	Settings>Network>General	20			
		7.4.3	Settings>Network>LAN	21			
		7.4.4	Settings>Network> WiFi	22			
		7.4.5	Settings>Measurement	22			
		7.4.6	Settings>Magnification	23			
		7.4.7	Settings>Image Format	23			
		7.4.8	Settings>Video	24			
		7.4.9	Settings>Storage	24			
		7.4.10	Settings>Files	25			
		7.4.11	Settings>Time	25			
		7.4.12	Settings>Language	25			
		7.4.13	Settings>ISP	26			
		7.4.14	Settings>EDF	26			
		7.4.15	Settings>Miscellaneous	27			
8	Sample Photos Captured with X5CAM4K Series Camera						
9	Contacting Customer Service						

### 1 X5CAM4K Series Camera Application



Figure 1 The X5CAM4K Series Camera(Cubic and Flat Shape)

The X5CAM4K series camera is intended for acquisition of digital images from stereo microscopes, biological microscopes, or online interactive teaching. The basic characteristic is listed as below:

- Sony STARVIS 2 back-illuminated CMOS sensor
- 4K HDMI/ NETWORK/ USB multiple video synchronous outputs
- 4K/1080P auto switching according to monitor resolution
- Support 4K 60fps low delay HDMI output mode, with an average delay of 40ms
- SD card/USB flash drive for captured image and video storage, support local preview and playback
- New browsing function, providing rich file operation functions, image to image comparison, image to real-time video comparison, Multi image EDF and other functions
- Excellent ISP with local tone mapping and 3D denoising
- Provide real-time video EDF function and real-time video WDR output function
- Provide two sets of default ISP parameters for biological microscope and stereo microscope
- Embedded XCamView for the control of the camera and image processing, supporting automatic edge finding and measurement functions
- ToupView/ToupLite software for PC
- iOS/Android applications for smart phones or tablets

## 2 X5CAM4K Series Camera Datasheet and Functions (2)

Order Code	Sensor & Size(mm)	Pixel(µm)	G Sensitivity Dark Signal	Sensor Output (FPS/Resolution)	Binning	Exposure(ms)
X5CAM4K8MPA	Sony IMX678(C) 1/1.8"(7.68x4.32)	2.0x2.0	3541mv with 1/30s 0.15mv with 1/30s	60@3840*2160	1x1	0.019~1000
X5CAM4K8MPB	Sony IMX585(C) 1/1.2"(11.14x6.26)	2.9x2.9	5970mv with 1/30s 0.13mv with 1/30s	60@3840*2160	1x1	0.048~1000

Camera Model	Video Saving(FPS/Resolution)	HDMI2.0(FPS/Resolution)	USB3.0(FPS/Resolution)	NETWORK(FPS/Resolution)
X5CAM4K8MPA	60@3840*2160 60@1920*1080	60@3840*2160 60@1920*1080	30@3840*2160 45@2688*1512 60@1920*1080	30@3840*2160 60@1920*1080 60@1280*720
X5CAM4K8MPB	60@3840*2160 60@1920*1080	60@3840*2160 60@1920*1080	30@3840*2160 45@2688*1512 60@1920*1080	30@3840*2160 60@1920*1080 60@1280*720



Figure 2 Available Ports on the Back Panel of the Camera Body(Cubic and Flat Shape)

Interface or Button	Function Description
USB Mouse	Connect USB mouse for easy operation with embedded XCamView software(Flat shape without labelling)
USB3.0	Connect USB flash drive to save pictures and videos Connect 5G WiFi module to transfer video wirelessly in real time Connect USB microphone for audio and video recording
USB Video	Connect PC or other host device to realize video image transmission
HDMI	Comply with HDMI2.0 standard. 4K/1080P format video output and supporting automatic switch between 4K and 1080P format according to the connected monitors
LAN	LAN port to connect router and switch to transfer video
SD	SD card slot, comply with SDIO3.0 standard and SD card could be inserted for video and images saving
ON/OFF	Power switch
LED	LED status indicator
DC12V	Power adapter connection (12V/1A)
Video Output Interface	Function Description
HDMI Interface	Comply with HDMI2.0 standard;60fps@4K or 60fps@1080P
LAN Interface	Support real time resolution switching(4K/1080P/720P) H264 encoded video DHCP configuration or manual configuration Unicast/multicast configuration
WiFi Interface	Connecting 5G WiFi adapter (USB3.0 slot) in AP/STA mode
USB Video Interface	Connecting USB Video port of PC for video transfer H264/MJPEG format video
Other Function	Function Description
Video Saving	Video format: 8M(3840*2160) H264/H265 encoded MP4 file Video saving frame rate: 60fps
Image Capture	8M (3840*2160) JPEG/TIFF image in SD card or USB flash drive (Default SD card priority, priority can be modified in settings)
Measurement Saving	Measurement information saved in different layer with image content Measurement information is saved together with image content in burn in mode
ISP	Exposure(Automatic / Manual Exposure) / Gain, White Balance(Manual / Automatic / ROI Mode), Sharpening, 3D Denoise, Saturation Adjustment, Gamma Adjustment, Contrast Adjustment, Brightness Adjustment, Hue Adjustment, Color to Gray, 50HZ/60HZ Anti-flicker Function
Image Operation	Zoom In/Zoom Out(Up to 10X), Mirror/Flip, Freeze, EDF, Cross Line, Overlay, PIP, Browser(including Picture Browsing, Video Playback, Video Compare, Picture Compare, EDF, Image Processing), Measurement Function
Embedded RTC(Optional)	To support accurate time on board

Restore Factory Settings	Restore camera parameters to its factory status				
Multiple Language Support	English / Simplified Chinese / Traditional Chinese / Korean / Thailand / French / German / Spanish / Japanese / Italian / Russian / Dutch / Portuguese				
	Software Environment under Network/USB Video Output				
White Balance	Auto White Balance				
Color Technique	Ultra-Fine Color Engine				
Capture/Control SDK	Windows/Linux/macOS/Android Multiple Platform SDK(Native C/C++, C#/VB.NET, Python, Java, DirectShow, Twain, etc)				
Recording System	Still Picture or Movie				
Operating System	Microsoft <sup>®</sup> Windows <sup>®</sup> XP / Vista / 7 / 8 / 8.1 / 10 / 11(32 & 64 bit) OSx(Mac OS X) Linux				
PC Requirements	CPU: Equal to Intel Core2 2.8GHz or Higher				
	Memory: 4GB or More				
	Ethernet Port: RJ45 Ethernet Port				
	Display:19" or Larger				
	CD-ROM				
	Operating Environment				
Operating Temperature (in Centidegree)	-10°~ 50°				
Storage Temperature (in Centidegree)	-20°~ 60°				
Operating Humidity	30~80%RH				
Storage Humidity	10~60%RH				
Power Supply	DC 12V/1A Adapter				

## 3 Dimension of X5CAM4K Series Camera



Figure 3 Dimension of X5CAM4K Series(Cubic and Flat Shape)

## 4 X5CAM4K Series Camera Packing Information



Figure 4 X5CAM4K Series Camera Packing Information(Cubic and Flat Shape)

	Standard Packing List			
Α	Gift box : L:25.5cm W:17.0cm H:9.0cm (1pcs, 1.7Kg/ box)			
В	X5CAM4K Camera(One of the two different shapes)			
С	Power Adapter: Input: AC 100–240V 50Hz/60Hz, Output: DC 12V 1A American standard: Model: POWER-U-12V1A(MSA-C1000IC12.0-12W-US): UL/CE/FCC European standard: Model: POWER-E-12V1A(MSA-C10001C12.0-12W-DE): UL/CE/FCC EMI standard: FCC Part 15 Subpart B EMS standard: EN61000-4-2,3,4,5,6			
D	USB Mouse			
Е	HDMI Cable			
F	USB3.0 A male to A male gold-plated connectors cable /2.0m			

G	CD (Driver & utilities software, Ø12cm)				
	Optional Accessory				
Η	SD Card(16G or above; S	peed: class 10)			
Ι	USB flash drive				
J	Adjustable lens adapter	C-Mount to Dia.23.2mm Eyepiece Tube (Please choose 1 of them for your microscope)	108001/AMA037 108002/AMA050 108003/AMA075		
K	Fixed lens adapter	C-Mount to Dia.23.2mm Eyepiece Tube (Please choose 1 of them for your microscope)	108005/FMA037 108006/FMA050 108007/FMA075		
	Note: For J and K optional items, please specify your camera type(C-mount, microscope camera or telescope camera), ToupTek engineer will help you to determine the right microscope or telescope camera adapter for your application;				
L	108015(Dia.23.2mm to 3	0.0mm Ring)/Adapter rings for 30mm	eyepiece tube		
Μ	108016(Dia.23.2mm to 30.5mm Ring)/ Adapter rings for 30.5mm eyepiece tube				
N	Calibration kit         106011/TS-M1(X=0.01mm/100Div.);           106012/TS-M2(X, Y=0.01mm/100Div.);         106012/TS-M7(X=0.01mm/100Div.);           106013/TS-M7(X=0.01mm/100Div.)         100013/TS-M7(X=0.01mm/100Div.);				
0	USB WiFi adapter				
Р	Ethernet cable				

## 5 Software and App

The software or the APP can be downloaded from the following link: Windows: <u>https://www.touptekphotonics.com/download/</u> Linux & macOS: <u>https://www.touptekphotonics.com/download/</u> iOS: <u>https://itunes.apple.com/us/app/toupview/id911644970</u> Android: <u>https://play.google.com/store/apps/details?id=com.touptek.tpview</u>

#### 6 X5CAM4K Series Camera Configurations

You can use the X5CAM4K series camera in 5 different ways. Each application requires different hardware environment.

#### 6.1 Camera working standalone with built-in XCamView software

For this application, apart from the microscope, you only need an HDMI monitor, the supplied USB mouse, and the camera embedded XCamView software. A computer or a network connection is not required to operate the camera in this application. The steps to start the camera are listed as below:



Figure 5 X5CAM4K Series Camera with the HDMI Monitor

Connect the camera to a HDMI monitor using the HDMI cable;



Insert the supplied USB mouse to the camera's USB Mouse port;



Insert the supplied SD card/USB flash drive into the X5CAM4K series camera SD card slot/USB3.0 slot;





Connect the camera to the power adapter and turn it on;



Turn on the monitor and view the video in the XCamView software. Move the mouse to the left, top or bottom of the XCamView UI, different control panel or toolbar will pop up and users could operate with the mouse at ease.



Figure 6 XCamView And X5CAM4K Series Camera in HDMI Mode

#### 6.2 Connecting camera to computers with USB3.0 port

For Windows user (Windows XP (32bit), Windows 7/8/10/11 (32/64 bit)), please use ToupView.

For macOS and Linux user (macOS 10.10 or above or Linux distributions with kernel 2.6.27 or higher), please use ToupLite. The steps to start the camera are listed below:

Start the camera according to Sec. 6.1. After the camera is running, connect camera to computer with USB cable. Please use "USB Video" slot, The upper left corner of the HDMI graphics interface displays "USB3.0 Mode" or "USB2.0 Mode", indicating that a connection has been established with the PC.



Install ToupView/ToupLite on your PC or install ToupView App on the mobile device; Run the software ToupView/ToupLite, clicking the camera name in the Camera List group to start the live video as shown in Figure 7.



Figure 7 ToupView and X5CAM4K Series Camera in USB Mode

#### 6.3 Camera working in WiFi mode (AP mode)

Please make sure your PC is WiFi enabled.



Figure 8 The PC or Mobile Device Connect to the Camera through WiFi

For Windows user (Windows XP (32bit), Windows 7/8/10/10/11 (32/64 bit)), please use ToupView.

For macOS and Linux user (macOS 10.10 or above or Linux distributions with kernel 2.6.27 or higher), please use ToupLite. When connecting the camera with a mobile device, the free ToupView App is required. Just make sure that the mobile device uses iOS 11 or higher/Android 5.1 or higher operating systems.

The steps to start the camera are listed below:

Start the camera according to Sec. 6.1. After the camera is running, move the mouse to the bottom of the GUI and clicking the  $\aleph$  button on the Synthesis Camera Control Toolbar at the bottom of the video window, a small window called Settings will pop up as shown below. Click Network> WiFi property page and choose the AP in the WiFi Mode edit box(The factory default configuration is AP mode ).

8	Settings	×
Network	General LAN WiFi	
Measurement Magnification Image Format Video Storage Files Time Language ISP EDF Miscellaneous	WiFi Mode: AP • Channel: 36 • Password: 12345678	
		Close Apply

Plug the USB WiFi adapter into the camera's USB3.0 port, the upper left corner of the HDMI graphics interface will display "AP mode";



Install ToupView/ToupLite on your PC or install ToupView App on the mobile device, connect the PC or mobile device to the camera's WiFi AP point; The network name (SSID) and the WiFi password (The default one is 12345678) can be found on the camera's Setting>Network> WiFi page in AP mode.



Start ToupView/ToupLite software or ToupView App and check the configuration. Normally, the active X5CAM4K series cameras will be automatically recognized. The live image of each camera is shown in Figure 9. For the display, the Camera List group is used in ToupView/ToupLite software, and the Camera Thumbnail is used in ToupView App.



Figure 9 ToupView and X5CAM4K Series Camera in WiFi AP Mode

#### 6.4 Connecting camera to the PC with LAN port

This application uses the camera as the network camera. User must configure the IP of the camera and PC manually and ensure their IP addresses in the same net. The subnet mask and gateway of the camera and PC must be the same.



Figure 10 Connecting the X5CAM4K Series Camera with Ethernet Cable to the PC

Start the camera according to Sec. 6.1 after the camera is running, clicking 💥 button on the Synthesis Camera Control Toolbar at the bottom of the video window(See Figure 6), a small window called Settings will pop up as shown below on the left side, clicking LAN property page, uncheck the DHCP item. Input IP Address, Subnet Mask and Default Gateway for the camera. Designate Internet Protocol Version 4 (TCP/IPv4) Settings page's IP address on the PC with similar configuration as shown below on the right side but with different IP address.

Intern	Internet 协议版本 4 (TCP/IPv-
Gene	General
You this for II S C C P P A	You can get IP settings assign this capability. Otherwise, you for the appropriate IP setting: O Obtain an IP address aut O Use the following IP add IP address: Subnet mask: Default gateway: O Obtain DNS server addre O Obtain DNS server: Alternate DNS server:

Figure 11 Configure the X5CAM4K Series Camera IP

Figure 12 Configure the PC's IP

After the above configurations are finished, user can connect the X5CAM4K series camera to the computer through the Ethernet cable as shown below:

Connect the LAN port with the Ethernet cable to the PC's network port, the upper left corner of the HDMI graphics interface will display IP address;



Insert the supplied SD card/USB flash drive into the X5CAM4K series camera's SD card slot/USB3.0 slot;



Install ToupView/ToupLite on your PC or install ToupView App on the mobile device; Run the software ToupView/ToupLite, clicking the camera name in the camera list starts the live video as shown in Figure 9.

6.5 Connecting multi-cameras to the router through the LAN port/WiFi STA mode for the network application

In LAN/ WiFi STA mode, the camera connects to the router by LAN port/ WiFi STA mode. If a router with LAN/ WiFi capability is used, users could connect the router with Ethernet cable/ WiFi to control the camera.



Figure 13 Multi X5CAM4K Series Cameras Connecting to the Router through the LAN Port/ WiFi Style

The connection and configuration are just the same as in Sec.6.1 or Sec. 6.4. But here, users need to check DHCP. If Multicast is disabled or is not supported, users should only select Unicast. If Multicast is supported by the network, users could select Multicast to achieve a better performance, especially in the case that multi-users connecting to the same camera. In addition, please guarantee that the broadcasting function is enabled in the network.

Active X5CAM4K series camera is recognized by ToupView/ToupLite software or ToupView App and they are displayed as a camera list or thumbnail in the software or app as shown in Figure 7.

*	Settings	×
Network       Measurement       Magnification       Image Format       Video       Storage       Files       Time       Language       ISP       EDF       Miscellaneous	Settings	×
		Close Apply

Or start the camera according to Sec. 6.1. After the camera is running, move the mouse to the bottom of the video window and clicking the  $\bigotimes$  button on the Synthesis Camera Control Toolbar at the bottom of the video window, a small window called Settings will pop up as shown below. Clicking Network> WiFi property page and choosing the STA in the WiFi Mode edit box(The factory default configuration is AP mode ). Input the to be connected router's SSID and Password as shown below:

*	Settings	×
Xetwork       Measurement       Magnification       Image Format       Video       Storage       Files       Time       Language       ISP       EDF       Miscellaneous	General LAN WiFi WiFi Mode: STA SSID: Password:	×
		Close Apply

Install ToupView /ToupLite software on your PC. Alternatively, install the free ToupView App on the mobile device;

Plug the Ethernet cable into the camera's LAN port and the other end to the PC (for those connected to router with LAN Port), the upper left corner of the HDMI graphics interface will display IP address;



Or plug the USB WiFi adapter into the camera's USB3.0 port(for those connected to router with WiFi STA mode), the upper left corner of the HDMI graphics interface will display "STA Mode";



Finally, as shown below, 2 X5CAM4K series cameras are connected to the router with LAN cable and 2 X5CAM4K series cameras are connected to the same router with WiFi STA mode (The number of the cameras, the connection mode (LAN or WiFi STA) connected to the router are determined by the router performance).



Make sure that your PC or your mobile device is connected to the LAN or WiFi of the router; Start ToupView/ToupLite software or ToupView App and check the configuration. Normally, active X5CAM4K series cameras are automatically recognized. The live image of each camera is displayed. For the display, Camera List group is used in ToupView/ToupLite software, and Camera Thumbnail is used in ToupView App; Select theX5CAM4K series camera you are interested in. To do so, double click the camera's name in Camera List tool window if you use ToupView /ToupLite software; If you use ToupView App, tap the camera's thumbnail in Camera List page(See Figure 14)

#### About the routers/switches

It is suggested that routers/switches supporting WiFi 5G should be selected to achieve better wireless connection experience.



Figure 14 ToupView and X5CAM4K Series Camera in LAN port/WiFi STA mode

## 7 Brief Introduction of X5CAM4K UI and Its Functions

#### 7.1 XCamView UI

The X5CAM4K UI shown in Figure 6 includes a Camera Control Panel on the left of the video window, a Measurement Toolbar on the top of the video window and a Synthesis Camera Control Toolbar on the bottom of the video window.

	Notes
1	To show the Camera Control Panel, move your mouse to the left or right of the video window. See Sec.7.2 for details
	Move the mouse cursor to the top of the video window, a Measurement Toolbar will pop up for calibration and measurement operations. When
	user left-clicks the Float/Fixed button 🖈 on the Measurement Toolbar, the Measurement Toolbar will be fixed. In this case the Camera
	Control Panel will not pop up automatically even if users move mouse cursor to left or right side of the video window. Only when user left-
2	clicks the 🗙 button on the Measurement Toolbar to exit from measuring procedure will they be able to do other operations on the Camera
	Control Panel, or the Synthesis Camera Control Toolbar. During the measuring process, when a specific measuring object is selected, an
i.	Object Location & Attributes Control Bar 🔗 🦞 🌏 ≽ 🐥 🧰 will appear for changing location and properties of the selected object.
	See Sec. 7.3 for details.
3	When users move mouse cursor to the bottom of the video window, the Synthesis Camera Control Toolbar will pop up automatically.
	(+) (-) / (-) (-) (-) (-) (-) (-) (-) (-) (-) (-)

#### 7.2 The camera control panel on the left or right side of the video window

The Camera Control Panel controls the camera to achieve the best video or image quality according to the specific applications; It will pop up automatically when the mouse cursor is moved to the left or right side of the video window (in measurement status, the Camera Control Panel will not pop up. The Camera Control Panel will only pop up when the measurement process is finished or terminated while user's cursor on the left edge of the video window). Left-clicking button to achieve Display/Auto Hide switch of the Camera Control Panel.

Camera Control Panel	Function	<b>Function Description</b>
	Snap	Capture image and save it to the SD card or USB flash drive
	Record	Record video and save it to the SD card or USB flash drive
	Auto Exposure	When Auto Exposure is checked, the system will automatically adjust exposure time and gain according to the value of exposure compensation
	Exposure Compensation	Available when Auto Exposure is checked. Slide to left or right to adjust Exposure Compensation according to the current video brightness to achieve proper brightness value
Camera Control Panel	Exposure Time	Available when Auto Exposure is unchecked. Slide to left or right to reduce or increase exposure time, adjusting brightness of the video
	Gain	Adjust Gain to reduce or increase brightness of video. The Noise will be reduced or increased accordingly
Snap Record	Red	Slide to left or right to decrease or increase the proportion of Red in RGB on video
☑ Auto Exposure: Exposure Compensation: 6	Green	Slide to left or right to decrease or increase the proportion of Green in RGB on video
Exposure Time: 13ms	Blue	Slide to left or right to decrease or increase the proportion of Blue in RGB on the video
Gain: 0	Auto	White Balance adjustment according to the window video every time the button is clicked
White Balance:	Manual	Adjust the Red, Green or Blue item to set the video White Balance
Red: 557	ROI	Check the ROI item will display a red ROI rectangle on the video window, drag it to the interested area will perform the White Balance according to the area video data
Blue: 372	One Push	Perform a global white balance based on image conditions
One Push	Sharpness	Adjust Sharpness level of the video
Sharpness : 40	Denoise	Slide left or right to Denoise the video
Saturation: 50	Saturation	Adjust Saturation level of the video
Gamma: 10 Contrast: 50	Gamma	Adjust Gamma level of the video. Slide to the right side to increase Gamma and to the left to decrease Gamma.
Brightness: 50	Contrast	Adjust Contrast level of the video. Slide to the right side to increase Contrast and to the left to decrease Contrast.
O     O     AC(50Hz) O     AC(60Hz)	Brightness	Adjust Brightness level of the video. Slide to the right side to increase Brightness and to the left to decrease Brightness.
Scence: Stereo •	Hue	Adjust Hue level of the video. Slide to the right side to increase Hue and to the left to decrease Hue.
	DC	For DC illumination, there will be no fluctuation in light source so no need for compensating light flickering
	AC(50HZ)	Check AC(50HZ) to eliminate flickering caused by 50Hz illumination
	AC(60HZ)	Check AC(60HZ) to eliminate flickering caused by 60Hz illumination
	Scence	Select different default parameters according to the type of microscope
	Default	Restore all the settings in the Camera Control Panel to default values Right click to select different default parameters according to the type of microscope

#### 7.3 The Measurement Toolbar on top of the video window

The Measurement Toolbar will pop up when moving mouse cursor to any place near the upper edge of the video window. Here is the introduction of the various functions on the Measurement Toolbar:

Icon	Function
-the	Float/ Fix switch of the Measurement Toolbar
✓ Visible	Show / Hide Measurement Objects
Pixel -	Select the desired Measurement Unit
NA	Select Magnification for Measurement after Calibration
×	Object Select
A	Angle
/\	4 Points Angle
•	Point(Point Counter)
/	Arbitrary Line
$\checkmark$	3 Points Line
/	Horizontal Line
	Vertical Line
$\times$	3 Points Vertical Line
11	Parallel
	Rectangle
$\diamond$	3 Points Rectangle
0	Ellipse
$\bigcirc$	5 Points Ellipse
Θ	Circle
0	3 Points Circle
0	Annulus
$\odot$	3 Points Annulus
P	Two Circles and its Center Distance
Ø	3 Points Two Circles and its Center Distance
0	Arc
T	Text
	Polygon
S	Curve
um	Scale Bar
7	Arrow
83	Execute Calibration to determine the corresponding relation between magnification and resolution, which will establish the corresponding relationship between measurement unit and the sensor pixel size. Calibration needs to be done with the help of a micrometer. For detailed steps of carrying out Calibration please refer to ToupView help manual.
	Auto Measurement: Two Points Parallel, Circle Detect, Annulus Detect, Rectangle Detect, Polygon
	Export the Measurement information to CSV file(*.csv)
(Y	Measurement Setup
Î	Delete all the measurement objects
×	Exit from Measurement mode
	When the measurement ends, left-click on a single measuring object and the Object Location & Properties Control Bar will show up. User could move the object by dragging the object with the mouse. But more accurate movement could be done with the control bar. The icons on the control bar mean Move Left, Move Right, Move Up, Move Down, Color Adjustment and Delete.

Note:

1) When user left-clicks Display/Hide button 🔌 on Measurement Toolbar, Measurement Toolbar will be fixed. In this case Camera Control Panel will not pop up automatically even if moving the mouse cursor to the left edge of the

video window. Only when user left-click the 🗙 button on Measurement Toolbar to exit from the measurement mode will they be able to doing other operations on Camera Control Panel or Synthesis Camera Control Toolbar.

2) When a specific Measurement Object is selected during the measurement process, Object Location & Attributes Control Bar &  $\forall \leq \&$  s s will appear for changing the object location and properties of the selected objects.

7.4 Icons and functions of the Synthesis Camera Control Toolbar at the bottom of the video window

Figure 16 The Synthesis Camera Control Toolbar on the Bottom of the Video Window

Icon	Function	Icon	Function
$\oplus$	Zoom In the Video Window	$\Theta$	Zoom Out the Video Window
	Horizontal Flip		Vertical Flip
(C+G)	Color/gray		Video Freeze
EDF	EDF	#	Display Cross Line
	Image Overlay	PIP	PIP
67	Browse images and videos in the SD Card	X	Settings
i	Check the Version of XCamView		

The 🗊 Browsing function, for detailed introduction, please refer to Section 7.4.1.

The 💥 Setting function, for detailed introduction, please refer to Sections 7.4.2 to 7.4.14.

#### 7.4.1 Browse

Clicking the 🗊 to browse the dxf, images, videos, and other files saved on the SD card or USB flash drive, as shown in the following figure.



#### Figure 17 Browsing UI

There are two browsing modes: List mode and Thumb mode. The default is Thumb mode.

Right click on an empty area to create a new folder.

Right click on an image file to Copy, Cut, Rename, Delete, Video Compare, and view detailed information(Details). Clicking on a thumb to select the 1<sup>st</sup> image, and clicking on another thumb to select the 2<sup>nd</sup> image (or selecting 2 images with frame), then clicking the right mouse button to bring up the context menu and select Picture Compare to analyze and compare the two images. Clicking on a thumb to select 2~5 (or box select 2~5) pictures focusing on different targets in the same scene, you can perform depth of field compositing on the selected pictures.

Right click on a video file to Copy, Cut, Rename, Delete, Video Compare, and view detailed information(Details).



Figure 18 Image Processing

Double-click the thumbnail of the picture with the left mouse button to open the picture, and then right-click the picture to Gray Scale, Invert, Highlights, Binary, Filter Color, Extract Color, Auto Level, Auto Contrast, Histogram, Histogram Equalization, Flip, and other image processing functions, and then after the processing is completed, you can choose reset to revert back to the original picture, and also you can choose save or save as in the lower sidebar of the picture. The description of each function is as follows:

Gray Scale	Choose Gray Scale command to convert a color image to a Gray Scale image
Invert	Choose Invert command to reverse the pixel values of the active image
Highlights	Choose Hightlights command to adjust the Hightlight parts of the images
Binary	Binary is a kind of gray level process. If the gray of the pixel is greater than the given threshold, the pixel's color will be changed into white. Otherwise, the pixel's color will be changed into black
Filter Color	Choose Filter Color command to filter a special color channel from a color image. Select either Red, or Green, or Blue color to filter. For every pixel, if select Red color to filter, only information about the Red channel will be discarded, the Green and Blue information will remain there.
Extract Color	Choose Extract Color command to extract a special color channel from a color image. Select either Red or Green, or Blue color to extract. For every pixel, if selecting Red color to extract, only information about the Red channel will be kept, the Green and Blue information will be discarded.
Auto Level	The Auto Level command moves the level's sliders automatically to set highlight and shadow. It defines the lightest and darkest pixels in each color channel as white and black and then redistributes the pixels' color values proportionately
Auto Contrast	The Auto Contrast command automatically adjusts the overall contrastin an RGB image
Histogram	Used to show the distribution of brightness, R, G, B of an image over an image
Histogram Equalization	Used to improved image contrast
Flip	Flip image Horizontally/Vertically

#### 7.4.2 Settings>Network>General

Â	Settings	×
Network Measurement Magnification Image Format	General LAN WiFi Name: X5CAM4K8MPA	
Video Storage Files Time Language		
ISP EDF Miscellaneous		
		Close Apply

Figure 19 Comprehensive Network General Settings Page

Name	The current camera name recognized as the network name

#### 7.4.3 Settings>Network>LAN

8	Settings	×
Network	General LAN WiFi	
Measurement Magnification Image Format Video Sitorage Files Time Language ISP EDF Miscellaneous	☑ DHCP       ● Unicast         IP Address:	
		Close Apply

Figure 20 Comprehensive Network LAN Settings Page

DHCP	Dynamic host control protocol allows DHCP server to automatically assign IP information to the camera. Only in Sec 6.4 LAN networking this item should be checked, so that cameras can automatically get IP information from routers/switches to facilitate networking operation;
Unicast/Multicast	By default, unicast function is used. Only in Sec 6.4 networking environment, when the router/switch has multicast function, camera can switch to multicast mode, which can save the network bandwidth consumed by the camera and facilitate the connection of more cameras in the same network;
IP Address	Every machine on a network has a unique identifier. Just as you would address a letter to send in the mail, computers use the unique identifier to send data to specific computers on a network. Most networks today, including all computers on the Internet, use the TCP/IP protocol as the standard for how to communicate on the network. In the TCP/IP protocol, the unique identifier for a computer is called IP address. There are two standards for IP address: IP Version 4 (IPv4) and IP Version 6 (IPv6). All computers with IP addresses have an IPv4 address, and many are starting to use the new IPv6 address system as well. Users must manually configure their IP addresses on the camera side and computer side. The IP addresses set on the camera side and computer side should be in the same network segment. The specific settings are shown Figure 21. It's usually a private address. Private address is a non-registered address used exclusively within an organization. The internal private addresses retained are listed below: Class A 10.0.0-10.255.255; Class B 172.16.0-172.31.255.255; Class C 192.168.0-192.168.255.255. The suggested IP address is Class C.
Subnet Mask	Subnet Mask is used to distinguish network domain from host domain in 32-bit IP address;
Default Gateway	A default gateway allows computers on a network to communicate with computers on another network. Without it, the network is isolated from the outside. Basically, computers send data that is bound for other networks (one that does not belong to its local IP range) through the default gateway; Network administrators configure the computer's routing capability with an IP range's starting address as the default gateway and point all clients to that IP address.

Uncheck the DHCP and select the Unicast item, user still need to set the IP address, Subnet mask and Default Gateway as shown below:

8	Settings	×
Network	General LAN WiFi	
Measurement Magnification Image Format Video Storage Files Time Language ISP EDF Miscellaneous	□ DHCP       ● Unicast       ○ Multicast         IP Address:       192       168       100       2         Subnet Mask:       255       255       25       0         Default Gateway:       192       168       100       1	
		Close Apply

Figure 21 Manual DHCP and Unicast

Uncheck the DHCP and select the Multicast item, user still need to set the IP address, Subnet Mask and Default Gateway as shown below:

8	Settings	×
Network	General LAN WiFi	
Measurement Magnification Image Format Video Sibrage Files Filme aanguage ISP EDF Miscellaneous	□ DHCP         ○ Unicast         ● Multicast           IP Address:         192, 168, 100, 2           Subnet Mask:         255, 255, 255, 0           Default Gateway:         192, 168, 100, 1	

Figure 22 Manual DHCP and Multicast

#### 7.4.4 Settings>Network> WiFi

8	Settings	×
Network Measurement Magnification Image Format Video Storage Files Time Language ISP EDF Miscellaneous	General LAN WiFi WiFi Mode: STA SSID: Password:	
	C	lose Apply

Figure 23 Network Setup

Wi-Fi Mode	AP/STA mode to select;
Channel/SSID	Channel for the AP mode and SSID for the STA mode. Here, the SSID is the router's SSID;
Password	Camera Password for the AP mode. Router Password for the STA mode

#### 7.4.5 Settings>Measurement

This page is used for the define of the Measurement Object properties.

8		Settings	×
Xetwork  Magnification  Image Format Video Storage Files Time	Global     Precision     Edge Detection     Font Size     Cursor     Miscellancous     Calibration     Line Width     -Color     * Angle	Settings The Calculation results keep 2 decimals Large Hide the label when moving the measurement object 2	
Language ISP EDF Miscellaneous	Line Width Color Label Type Point Line Width Color Label Type Line Line Width	2 ☑ Angle 2 ☑ Position 2	
	-Color Label Type	M Leneth	Default Close Apply

Figure 24 The Measurement Setup

	Precision	Used for setting digits behind the decimal point for measurement results;	
	Edge Detection	Select whether to enable the automatic edge search function and set the detection range;	
Global	Font Size	The font size of measurement data can be divided into three types: large, Middle, and Small;	
	Cursor	Select whether the cursor is a single crosshair and set the color of the single cross;	
	Miscellaneous	Whether to hide the label when moving the measurement objects;	
	Line Width	Used for defining width of the lines for calibration;	
Calibration	Color	Used for defining color of the lines for calibration;	
Cultoration	EndPoint	Type: Used for defining shape of the endpoints of lines for calibration: Null means no EndPoint, rectangle means rectangle type of endpoints. It makes alignment more easily;	
Point, Angle, Line, Horizontal Line, Vertical Line, Rectangle, Circle, Ellipse, Annulus, Two Circles, Polygon, Curve			
	Left-click the in along with the Measurement command mentioned above will unfold the corresponding attribute settings to set		
	the individual property of the Measurement Objects.		

#### 7.4.6 Settings>Magnification

This page's items are formed by the Measurement Toolbar's Calibration command.

ŝ.		Settings	)
letwork	Name	Resolution	Clear All
Aeasurement 1 42	K C	800000000.00	Delete
Agnification 2 10	X	400000000.00	Up
nage Format			Down
ïdeo			
torage			
iles			
ime			
anguage			
SP			
DF			
fiscellaneous			
			Close Apply

Figure 25 Comprehensive Magnification Settings Page

Name	Names such as 10X, 40X, 100X are based on magnification of the microscopes. For continuous zoom microscopes, ensure that the selected magnification coincides with the scale alignment line on the microscope zoom knob; Users could also edit the name of the magnification with other information, for example, microscope mode, users name, etc.
Resolution	Pixels per meter. Image device like microscopes have high Resolution value;
Clear All	Click the Clear All button will clear the calibrated magnifications;
Delete	Click Delete to delete the selected magnification;
Up	Select a row in the magnification and click Move Up to move up the currently selected magnification;
Down	Select a row in the magnification and click Move Down to move up the currently selected magnification;

#### 7.4.7 Settings>Image Format

8	Settings	×
Network Measurement Magnification Image Format Video Storage Files Time Language ISP EDF Miscellaneous	Settings         Image Format       Image Format         Image Fore	
	Close Ap	ply

#### Figure 26 Comprehensive Image Format Settings Page

Image Format	JPEG: The extension of JPEG file can get very high compression rate and display very rich and vivid images by removing redundant images and color data. In other words, it can get better image quality with the least disk space. If measurement objects are available, the measurement objects will be burned into the image and the measurement cannot be edited. TIFF: TIFF is a flexible bitmap format mainly used to store images including photos and artistic images.
Measurement Object Saving Method	Burn in Mode: The measurement objects are merged into the current image. User could not edit the measurement objects any more. This mode is not reversable. Layered Mode: The measurement objects are saved in different layer with current image data in the target file. User could edit the measurement objects in the target file with some software on the PC. This mode is reversable.

#### 7.4.8 Settings>Video

8	Settings		>
Vetwork     Measurement     Magnification     mage Format     //deo     Vetwork     V	Settings       Video Resolution       0 1920x1080       3840x2160       Video Encode       H264       H265       Playback       Fast Forward/Reverse Interval: 20 \$ seconds		
Aiscellaneous			
		Close	Apply

#### Figure 27 Comprehensive Setting of Video page

Video Resolution	Select a Video Resolution of 1920x1080 or 3840x2160;
Video Playback	Fast Forward/Reverse internal in second unite for Video Playback
Video Encode	Select the Video Encode format. Can be H264 or H265. Compared with H264, H265 has a higher H265 compression ratio which is primarily used to further reduce the design flow rate, in order to lower the cost of storage and transmission

#### 7.4.9 Settings>Storage

8		Settings	×
Network Measurement Magnification Image Format Video	Preferred Storage Device SD Card USB Flash Drive File System Format of the Storage I	Device	
Storage	SD Card	USB Flash Drive	
Files	© FAT32	O FAT32	
Time	ONTES	ONTES	
Language	O Unknown Status	O Unknown Status	
ISP			
EDF			
Miscellaneous			
			Close Apply

Figure 28 Comprehensive Setting of Storage Page

Preferred Storage	SD Card: Select it to save the video and image to the SD Card.
Page	USB Flash Drive: Select it to save the video and image to the USB Flash Drive.
	List the file system format of the current storage device
File System	FAT32: The file system of SD Card is FAT32. The maximum video file size of single file in FAT32 file system is 4G Bytes;
Format of the	exFAT: The file system of SD Card is exFAT. The maximum video file size of single file in FAT32 file system is 16E Bytes;
Storage Device	NTFS: The file system of SD Card is NTFS. The maximum video file size of single file is 2T Bytes.
_	Unknown Status: SD Card not detected or the file system is not identified;
Note: For USB Flas	h Drive. USB 3.0 interface is preferred.

## 7.4.10 Settings>Files

8		Settings	×
Network	Image File Name		
Measurement	<ul> <li>Auto</li> </ul>	○ Manual	
Magnification	Prefix: IMG		
Image Format	Video File Name		
Video	Auto	⊖ Manual	
Storage	Prefix: VID		
Files			
Time			
Language			
ISP			
EDF			
Miscellaneous			
			Close Apply

Figure 29 Comprehensive Setting of Files Name

Image or Video File Name Paradigm	Provide Auto or Manual naming paradigm for Image or Video file;
Auto	With specified name as the Prefix and XCamView will add digital after the Prefix for the Image or Video file;
Manual	A file dialog will pop up to enter the Image or Video file name for the captured Image or Video.

## 7.4.11 Settings>Time

8			Settings		×
★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★★	2023-08- TimeZone UTC+08: Year: 2023 Hour: 5	01 05:54:21  Month: 8 Minute: 54	Day:		×
		<b>C</b> '	20 T	C. 11 <sup>1</sup>	Close Apply

Time User can set Year, Month, Day, Hour, Minute and Second ital.in this page.

#### 7.4.12 Settings>Language

Network  Senglish  Measurement  Simplified Chinese (简体中文)  Magnification  Traditional Chinese (驚體中文)  Magnification  Traditional Chinese (驚體中文)  Korean (같'= ^-)  Thailand (n'ay Thu)  Thailand (n'ay Thu)  Files  Spanish (Español)  Files  Auguage  Russian (pyecknii)  EDF  Potuguese (Portuguès)  Miscellaneous  Summary Senglish  Simplified Chinese (Portuguès)  Summary Senglish  Summary Seng	8	Settings	×
Measurement Simplified Chinese (简体中文) Magnification Traditional Chinese (繁體中文) Image Format Ctorese (繁體中文) Thailand (การา Tus) Video French (Francais) Storage German (Deutsch) Files Spanish (Español) Time Japanese (日本語) Language Rational (Italiano) Russian (pyccxrili) ISP Dutch (Nederlands) EDF Portuguese (Português)	Network	English	
Magnification ( ) Traditional Chinese (繁體中文) ( Korean (한국어) ( Thailand (การา 1\mu) Video	Measurement	○ Simplified Chinese (简体中文)	
Image Format       (Norean (2 <sup>+2</sup> °1)         (Thailand (n'n' Thu)         Video       French (Francais)         Storage       German (Deutsch)         Files       Spanisk (Español)         Time       Japanese (日本語)         Canguage       Outh (Nederlands)         ISP       Dutch (Nederlands)         EDF       Portuguese (Portugués)	Magnification	○Traditional Chinese (繁體中文)	
Video     Initiatio (1/19/160)       OF French (Francais)       Storage     German (Deutsch)       Files     Spanish (Español)       Time     Japanese (日本語)       Initianio     Initiano (1/1000)       Cangunge     Outpanese (Portugués)       ISP     Ortuguese (Portugués)       Miscellaneous     Portuguese (Portugués)	Image Format	○Korean (안국어) ○ Theilerd (courd line)	
Storage       German (Deutsch)         Files       Spanish (Español)         Imaguage       Japanese (日本語)         Italian (italiano)       Russian (pyccwiti)         ISP       Duth(Nederlands)         EDF       Portuguese (Português)	Video	O French (Francais)	
Files     Opanish (Español)       Time     Japanese (日本語)       Cangunge     Otalian (taliano)       BEDF     Oportuguese (Portugués)	Storage	O German (Deutsch)	
Time     ○ Japanese (日本語)       Language     ○ Italian (italiano)       SEP     ○ Dutch (Nederlands)       EDF     ○ Portuguese (Português)	Files	O Spanish (Español)	
Language     O tatalati (tatalato)       O Russian (pyccxtii)       ISP     O Dutch (Nederlands)       EDF     O Portuguese (Português)       Miscellaneous	Time	○ Japanese (日本語)	
ISP OLuch (Vederlands) EDF Portuguese (Português) Miscellaneous	Language	O Russian (папано)	
EDF O Portuguese (Portugués) Miscellaneous	ISP	O Dutch (Nederlands)	
Miscellaneous	EDF	○ Portuguese (Português)	
	Miscellaneous		
Close Anniv		Close	hpply

#### Figure 31 Comprehensive Setting of Language Selection Setting Page

English	Set language of the whole software into English;
Simplified Chinese	Set language of the whole software into Simplified Chinese;
Traditional Chinese	Set language of the whole software into Traditional Chinese;
Korean:	Set language of the whole software into Korean;
Thailand	Set language of the whole software into Thailand;
French	Set language of the whole software into French;
German	Set language of the whole software into German;
Spanish	Set language of the whole software into Spanish;
Japanese	Set language of the whole software into Japanese;
Italian	Set language of the whole software into Italian;
Russian	Set language of the whole software into Russian;
Dutch	Set language of the whole software into Dutch;
Portuguese	Set language of the whole software into Portuguese;

#### 7.4.13 Settings>ISP

8	Settings		×
Network Measurement Magnification Image Format Video Storage Files Time Logguego	Auto Exposure Maximum exposure time: 13 ms Metering Mode O Centre Weighted Average Metering @ Evaluative Metering O Partial Metering O Spot Metering		
Language	WBROI		
EDF Miscellaneous	Color.		
		Close	Apply

## Figure 32 Comprehensive Setting of ISP Page

Auto Exposure	Define the maximum automatic exposure time;
Metering Mode	Select the Metering mode as the Central Weighted Average Metering, Evaluative Metering, Partial Metering, or Spot Metering;
WB ROI Color	Choosing the ROI rectangle line color and whether it is synchronized display as Camera Control Panel;
Clarity Factor	Select to display the clarity factor in the video window, otherwise the clarity factor will not be displayed;
Dark Enhance	Define the intensity value of dark enhancement;

#### 7.4.14 Settings>EDF

8			Settings	×
Network	Automatic Align	ment		
Measurement	() None	Shift		
Magnification Image Format	Sensitivity	@ Middle	⊖L av	
Video	Orligii	() Middle	Olow	
Storage	Window size O Large	⊖ Middle	Small	
Time	Description			
Language	Automatic align	ment: Solves offs	et issues of fused images, but slows down the fusion proces	SS.
ISP	image.	roves the detectio	n accuracy of depth of field, perhaps reduce the quality of f	used
EDF				
Miscellaneous				
			Cl	lose Apply

## Figure 33 Comprehensive Setting of EDF

Automatic Alignment	Optionally turn on auto-alignment when there is significant displacement between images;
Sensitivity	Select the sensitivity of EDF;
Window size	Select the window size for displaying real-time images during EDF;
Description	Automatic alignment: Solves offset issues of fused images, but slows down the fusion process. Sensitivity:Improves the detection accuracy of depth of field, perhaps reduce the quality of fused image.

# 7.4.15 Settings>Miscellaneous

8	Settings	×
Network Measurement Magnification Image Format Video Storage Files Time Language ISP EDF	Ruler Show Color: Keasurement Keasurement Support saving overlay information in Burn In Mode Grids Support saving grids information in Burn In Mode Monitor Working Mode Show	¢
Miscellaneous	Cursor Size: Middle • Camera Control Panel Display Location © Left ORight OBoth Camera Parameters Tmontt Export	

Figure 34 Comprehensive Miscellaneous Settings Page

Ruler	Select to display the ruler in the video window, otherwise not to display the ruler. You can choose the ruler color;
Measurement	Select to display the measurement toolbar in the video window, otherwise not to display the measurement toolbar;
Overlay	Select to support saving graphics overlay information in fusion mode, otherwise it will not support;
Grids	Select to support saving mesh information in fusion mode, otherwise not to support;
Monitor Working Mode	Select to display the Monitor Working Mode in the video window, otherwise the Monitor Working Mode will not be displayed;
Cursor	Choosing the Cursor size according to the screen resolution or personal preference
Camera Control Panel Display Location	Select the camera control panel to display on the left, right, or both sides of the HDMI interface;
Camera Parameters Import	Import the Camera Parameters from the SD Card or USB flash drive to use the previously exported Camera Parameters
Camera Parameters Export	Export the Camera Parameters to the SD Card or USB flash drive to use the previously exported Camera Parameters
Reset to factory defaults	Restore camera parameters to its factory status;

8 Sample Photos Captured with X5CAM4K Series Camera



Figure 35 Corn Root Tip.L.S Captured with X5CAM4K8MPA



Figure 36 Three Year Tilia Stem.C.S Captured with X5CAM4K8MPA



Figure 37 Corn Leaf Captured with X5CAM4K8MPA



Figure 38 Circuit Board Captured with X5CAM4K8MPA

## 9 Contacting Customer Service

Please contact your local distributor if you have any questions about the product.